

CNL Event Reporting 2018/2019 3rd Quarter (October, November, December)

Event Number	Title
P&NO-18-3601	Legacy Sources found that are greater than Exemption Quantity - not Properly Posted with Hazard Warning Signs
D&WM-18-3694	Potential Uptake for Facilities Decommissioning Worker
D&WM-18-3719	HWP - PG WWTP - Security Event
OPS-18-3801	WTC Patrol Area Radiation Monitors overdue for calibration
R&D-18-3884	Accessible whole body dose rate > 2.5 mrem/h in B375 Lab 258
D&WM-18-4109	HWP - PG WWTP - Bioreactor Spill of less than one cubic metre.
D&WM-18-4432	HWP - TSS -PSNE Suspected trespass into remediated area
R&D-18-4561	Only one qualified operator in the ZED-2 Facility with the Reactor in the Standby Condition.
R&D-18-4567	Legacy Radiation Device not in RP Registry - Omission of Leak Tests
OPS-18-4580	Universal Cell Exhaust Fan Environmental Monitoring Temporarily Impaired